

LabVIEW and National Instruments (NI) Hardware Tues, 14:00

Bob Strand

- I. Introduction - Why LabVIEW?
- II. Determinism
 - A. Program Counter (Instruction Pointer)
 1. Top-down Code
 2. Campbell model - Programmable Logic Controller
 - B. The real world doesn't always function in a timed loop. Many times real-world interaction is event-driven
 1. Counters (Anemometer, tipping-bucket rain gauge)
 2. Interrupt Service Routine (ISR) - 1989: Global Change data collection
 3. Fortunately, hardware manufacturers, Microsoft, Apple, etc., have handled much of the ISR coding necessary to write modern user applications.
- III. Determinism through data flow programming
 - A. Lucid - My first introduction
- IV. LabVIEW (Laboratory Virtual Instrument Engineering Workbench)
 - A. Data Flow Programming Language
 - B. Originally Developed for Apple Macintosh
 - C. Originally intended to facilitate the use of National Instruments hardware
 - D. Has gained popularity as a general use language
 - E. Interfaces available for a wide variety of third-party measurement and control devices
 - F. Graphical Programming
 - G. Graphical user interfaces
 - H. Environment includes robust debugging tools
 1. Graphical depiction of execution
 2. Breakpoints (including conditional)
 3. Single-step execution
 4. Ability to probe variable values (No print statements necessary)
 5. Code profiling
- V. National Instruments Hardware
 - A. Broad array of platforms hosting varying measurement and control configurations
 - B. Many platforms are modular
 - C. Fast and accurate
 - D. Greater acquisition rates allow for characterization and mitigation of noise
- VI. Introduction to phenotyping platforms of Kansas State University (Kevin)
 - A. System requirements: measure the canopy temperature, the canopy height, and canopy reflections at various spectral wavelengths
 - B. Two platforms:
 1. Phenotyping Mobile Unit (PheMU)
 2. Hand held phenotyping unit (PhenoCorn)

- C. PheMU design
 - 1. Three sets of sensors
 - a. GreenSeeker – NDVI
 - b. Crop Circle – NDVI
 - c. Ultrasonic sensor – canopy height
 - d. Infrared thermometer (IRT) – canopy temperature
 - 2. Two RTK GPS units
 - 3. A high-clearance sprayer vehicle
 - D. Data acquisition system design
 - 1. A laptop computer that runs a NI LabVIEW program to record the sensor outputs and save the data into text files
 - 2. Several types of hubs that extend the ports of the laptop to receive data from serial output sensors
 - 3. A NI 9207 analog input module that collects data from analog output sensors
 - E. Software development
 - 1. Front panels
 - a. Setup
 - b. Monitor
 - c. Raw Data
 - 2. Block diagrams
 - a. Sampling control module
 - b. Sensor control module
 - c. Data storage module
- VII. Demonstration of collection on PXI-Express chassis
- VIII. These tools (both Campbell and National Instruments) enable development, but... there are things to realize.
- A. There are learning curves associated with each of them
 - 1. Data flow programming requires a significant paradigm shift
 - B. Moving beyond simple acquisition and control can require machine-level expertise